

Weightron Bilanciai Ltd,
 Titan Works
 Bridge Way off Broombank Road,
 Chesterfield Trading Estate,
 Chesterfield S41 9QJ, UK
 Tel: 01246 260062
 Fax: 01246 260844
 www.weightron.com

Weightron Bilanciai Ltd,
 Unit 4b, Block 4
 Westpark, Chelston
 Wellington TA21 9AD, UK
 Tel: 01823 662355
 Fax: 01823 6621017
 www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the Calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Eco Waste Recycling Ltd	
Unit 5, Hullock Pit Hill Newham Road Baldock Hertfordshire SG7 5JX	
Location	Site

Machine Details

Cust ref	Weighbridge	
Make	Bilanciai	
Model	Flynet 50 Analogue	
Serial No	T270705672	
Capacity	50000	kg
e/d	20	kg
Working cap	44000	

Test Weights

Set No.	Cert No.	Date Cal
WB 50-71	TR12817	20/08/2026

Cert Interval	12 Months
Next Cert	Aug 2026

Tolerance +/-	kg
-	20
-	20
-	40
-	40
-	40
-	60

Comments

Make up weights used.



AS FOUND LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev
4000	4000	0	4000	0
10000	10000	0	10000	0
18000	17980	-20	17980	-20
23300	23280	-20	23280	-20
35200	35180	-20	35180	-20
44500	44460	-40		

DEFINITIVE LINEARITY & HYSTERESIS

Load kg	Loading	Dev	Unloading	Dev

Repeatability

Load kg	Ind	Ind	Ind
44500	44460	44460	44460

Repeatability

Load kg	Ind	Ind	Ind

Eccentric Load 7300

Reading	kg	Dev	Reading	kg	Dev
1	7300	0	2	7300	0
3	7280	-20	4	7280	-20
5	7300	0	6	7300	0
7	7300	0	8	7300	0
9			10		
11			12		

Eccentric Load

Reading	kg	Dev	Reading	kg	Dev
1			2		
3			4		
5			6		
7			8		
9			10		
11			12		

Signature:

Name: Keith Lampard

Date: 20/08/2025 08:59